

Test report - PMFT 1000 - Breath resistance / Valve test



Operator	Date	Time	Vesion
DeR	15.01.2025	12:47:49	1.12.2.0

Basic information

Operation parameter	
Script	Breath Resistance
Aerosol	Pressurized air

Test conditions	
Temperature	24.95 °C
rel. humidity	2.46 %

Mask information

Mask ID	SkyPro GL001A FFP2 Maske
Mask lot	FFP2-S241202
Mask type	FFP2 / x / x

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.91 l/min	40.76 Pa	70.00 Pa	✔ pass
Inhalation 2	94.42 l/min	131.01 Pa	240.00 Pa	✔ pass
Exhalation	159.37 l/min	233.42 Pa	300.00 Pa	✔ pass

Calibration valid until 09.04.2025