

Test report - PMFT 1000 - Breath resistance / Valve test



Operator	Date	Time	Vesion
MB	25.04.2024	15:25:43	1.12.2.0

Additional information

HCP Business GmbH
Melkerstieg 20, 21217 Seevetal, Deutschland
Frau Lu Freese
K-20574

Basic information

Operation parameter		Test conditions	
Script	Breath Resistance	Temperature	22.42 °C
Aerosol	Pressurized air	rel. humidity	11.30 %

Mask information

Mask ID	0145_20240425_152543
Mask lot	SkyPro GL001A 240102
Mask type	FFP2 / x / x

Result

Description	Main air	Differential pressure	Upper limit	Result
Inhalation 1	29.30 l/min	26.65 Pa	70.00 Pa	✓ pass
Inhalation 2	95.35 l/min	85.31 Pa	240.00 Pa	✓ pass
Exhalation	160.65 l/min	151.01 Pa	300.00 Pa	✓ pass

Calibration valid until 08.04.2025